



# CERTIFICATE OF ANALYSIS

**PRODUCT** CT1793S  
 CEFQUINOME 30µg (CEQ30) INTERVET GERMANY

**LOT NUMBER** 1498064

**EXPIRY DATE** 2015.11.30

**DATE OF MANUFACTURE** 2014.05.09

| Delivery/Customer information |            |
|-------------------------------|------------|
| Date Printed                  | 2014.08.20 |
| Delivery No.                  |            |
| Customer                      |            |
| Customer Order number         |            |

### Microbiological Performance

Assay (mcg/disc) 31.1

### Tested in accordance with current CLSI methodology

|  | Zone Size (mm) | Limits (mm) |
|--|----------------|-------------|
| <i>Staphylococcus aureus</i> ATCC@25923    | 32             | 25 - 33     |
| <i>Escherichia coli</i> ATCC@25922         | 29             | 28 - 36     |
| <i>Streptococcus pneumoniae</i> ATCC@49619 | 38             | 30 - 38     |

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed. Quality control ranges are internally derived based on historical data.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any application other than that specified in the current Oxoid Manual. The results reported were obtained at the time of release.

Lot Accepted, 2014.05.15

Carol MacBride  
 PRODUCT PERFORMANCE MANAGER

ATCC is a registered trade mark of the American Type Culture Collection.  
 NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.



**OXOID LIMITED**  
 Wade Road, Basingstoke, Hampshire RG24 8PW, England  
 oxoid@thermofisher.com www.oxoid.com

FDA Reg No. 8010096